



Precision Machinery Research & Development Center(Taiwan)

財團法人精密機械研究發展中心

Verification of Compliance

Automated Nucleic Acid Extractor

Model : MagCore® Super

Rating : 100-240Vac 50/60Hz 850W

Applicant / Holder :

RBC Bioscience Corp.

*15F., No.15, Qiaoh Rd., Zhonghe Dist., New Taipei City 235029,
Taiwan*

The verification of compliance refers to the above mentioned sample(s) complying to the relevant harmonized standard(s) are as follows :

IEC 61010-2-101:2018

EN IEC 61010-2-101:2022+A11:2022

This verification does not imply assessment of production of the product.
The decision principle does not adopt the uncertainty of measurement.

Verification No.: 113R0398-070

Issued Date: 2024-02-29

Expired Date: 2027-02-28

Approved by PMC



Precision Machinery Research & Development Center(Taiwan)

財團法人精密機械研究發展中心

Address : No.27, 37th Road, Taichung Industrial Park, Taichung, Taiwan, R.O.C.

Tel : 886-4-23599009 Fax : 886-4-23598847